


Lansoprazole



Method	LC/MS/MS
Matrix	Human Plasma
Anticoagulant	Sodium Heparin
Storage Temperature	-20 °C
LLOQ	0.25 ng/mL
ULOQ	300 ng/mL
Sample Volume	500 µL
Precision & Accuracy	
LLOQ QC	11.5 % CV; 4.0 % Bias
Low QC	7.8 % CV; 2.0 % Bias
Mid QC	7.3 % CV; 4.1% Bias
High QC	7.9 % CV; -6.9% Bias
Above Range QC (5x Dilution)	8.4 % CV; -13.5% Bias
Stability	
Short-Term (Bench-Top)	4 hrs
Long-Term	9.5 months at -20 °C
Freeze/Thaw Cycles	3
Comments	Validated in Presence of Omeprazole
Method Number	V1879P1